



ADVANCED TECHNOLOGIES IN MR FLUID MANAGEMENT

7457 Aloma Avenue, Suite 201  
Winter Park, Florida 32792  
Ph: 407.677.8022  
Fx: 407.677.5037  
www.iradimed.com

**DECLARATION OF CONFORMITY**

Application of Council Directive(s) Medical Device Directive 93/42/EC

Regulation(s) to which Conformity is Declared LVFS 2003:11

Standard(s) to which Conformity is Declared IEC 60601-1, IEC 60601-2-24, ISO 9919

Manufacturer's Name Iradimed Corporation

Manufacturer's Address 7457 Aloma Ave., Winter Park, FL 32792 USA

Importer's Name Refer to accompanying Packing Slip

Importer's Address Refer to accompanying Packing Slip

Type of Equipment MRIdium 3860 Infusion Pump System and accessories

Model No(s). 3860 Series and 1000 Series

Serial No(s). Refer to accompanying Packing Slip

Year of Manufacture Refer to accompanying Packing Slip

Certification Method(s) Annex II

Equipment Class Class IIa and IIb

MDD Conformity Assessed By Semko (0413)

I, the undersigned, hereby declare that the equipment specified above conforms to the above Directive(s), Regulation(s) and/or Standard(s).

Place Winter Park, FL

Date January 4, 2013

  
\_\_\_\_\_  
(Signature)  
Francis Casey  
\_\_\_\_\_  
(Full Name)  
Regulatory Affairs  
\_\_\_\_\_  
(Position)