

## Test Certificate

on the Effect of Third Party Products  
on MR Systems

### 1. Test Certificate, device designation and manufacturer

This is to certify that the following device

**MRidium 3850 MRI IV Pump**  
**tested with serial number 0025**

of the manufacturer:

**Iradimed Corp.**  
**7457 Aloma Avenue/Ste 201**  
**Winter Park FL 32792, USA**

does not restrict the safety and functioning of the Siemens AG, Medical Solutions Magnetic Resonance Systems described in Clause 2.

Possible functional restrictions of the MR Systems are given in Clause 4.

Possible adverse effects of the MR Systems on the device described above are explicitly not subject to the testing upon which this Certificate is based. Consequently, this Certificate does not imply that the above-mentioned device is not subject to disturbance by the MR System.

### 2. MAGNETOM Systems affected

Seq. no.	MR System or option (type designation)
1	Harmony 1.0 T
2	Symphony 1.5 T
3	Sonata 1.5 T
4	Avanto 1.5 T
5	Espreo 1.5 T
6	Trio 3.0 T

### 3. Scope

In connection with the MR application the MRidium 3850 MRI IV Pump is intended for the following use to which the Test Certificate refers:

- Infusion Pump for use in the MRI

### 4. Restrictions

During use of the MRidium 3850 MRI IV Pump there are the following restrictions of functions and/or application possibilities of the MR System, the systems, options, accessories:

- During an MR scan the AC Adapter (battery charger) must be switched off., otherwise the electromagnetic noise will be out of specification.

### 5. Warnings

When using the MRidium 3850 MRI IV Pump in connection with the MR Systems described in Clause 2 the following precautions must be observed:


- The AC Adapter (battery charger) contains ferromagnetic components and must be operated outside the 1000 Gauss line. A corresponding note must be included in the operator manual and on the product in the form of a warning notice.
- During an MR scan the AC Adapter (battery charger) must be switched off., otherwise the electromagnetic noise will be out of specification. A corresponding note must be included in the operator manual and on the product in the form of a warning notice.

### 6. Validity:

This Test Certificate is valid until revoked by Siemens AG, Medical Solutions.

Erlangen, 2006-02-14

Siemens Aktiengesellschaft

  
\_\_\_\_\_  
Dr. Söldner

  
\_\_\_\_\_  
Schumann